

12. 14-MEV NEUTRON ACTIVATION ANALYSIS OF SELECTED LEG 2 CORE SAMPLES

Activation Analysis Research Laboratory
Texas A & M University, College Station, Texas

SAMPLE PREPARATION

Each core sample was first crushed and mixed to permit packaging in small polyethylene vials. From each crushed core sample, two sub-samples weighing approximately 1 gram each were taken at random, weighed and placed in vials which were then heat sealed. Standards containing known amounts of oxygen, silicon, iron, magnesium, and aluminum were weighed and capsulated.

ANALYTICAL PROCEDURES

Each sample and standard was irradiated with 14 MeV neutrons for two minutes, counted for two and ten minutes after cooling times of ten minutes and two hours, respectively. A pneumatic sample transfer system was used to carry each sample to the irradiation terminal and then to a 3 inch X 3 inch sodium-iodide

(T1) detector coupled to a 400 channel pulse height analyzer for counting. A weighted least-squares computer program which utilizes the information contained in the two spectra was used to resolve the radioactivity from the activation products of silicon, aluminum, iron and magnesium and to compute the amount of each element present.

The oxygen content was determined by a separate procedure which consisted of five short irradiations and counts under standardized conditions. The observed activity was then compared to that of oxygen standards by a computer program written to accommodate these calculations.

The uncertainty in the analytical results is expressed here in absolute per cent.

TABLE 1.
Major Element Composition of Selected Core Samples From Leg 2

Sample Designation				Composition, Wt %				
Hole	Core	Section	Sampled at (cm)	O	Si	Fe	Al	Mg
8	2	2	36-37	47.9 ± 2.4	37.40 ± 1.90	2.68 ± 0.40	5.07 ± 0.76	1.60 ± 0.20
9	5	3	10-11	39.4 ± 2.0	12.70 ± 0.64	3.07 ± 0.46	3.99 ± 0.60	1.50 ± 0.20
9	7	4	10-11	35.5 ± 1.8	25.00 ± 1.20	6.67 ± 1.00	9.28 ± 1.40	2.30 ± 0.30
9	8	3	2-3	37.3 ± 1.9	28.00 ± 1.40	7.12 ± 1.10	10.07 ± 1.51	2.30 ± 0.30
9	8	6	3.5-5	35.8 ± 1.8	26.90 ± 1.30	7.63 ± 1.10	9.86 ± 1.48	2.20 ± 0.30
9	9	2	9.5-10.5	36.0 ± 1.8	26.50 ± 1.30	8.20 ± 1.20	10.62 ± 1.59	1.90 ± 0.30
9	10	3	10-12	33.8 ± 1.7	27.90 ± 1.40	6.83 ± 1.00	10.58 ± 1.59	2.30 ± 0.30
9	10	6	54-56	35.4 ± 1.8	28.40 ± 1.40	8.15 ± 1.20	10.72 ± 1.61	2.30 ± 0.30
9A	1	6	9-11	35.9 ± 1.8	26.50 ± 1.30	6.28 ± 0.94	9.59 ± 1.44	2.00 ± 0.30
9A	3	4	95-96	41.8 ± 2.1	38.40 ± 1.90	4.40 ± 0.66	3.67 ± 0.55	0.88 ± 0.13
9A	5	1	70	36.2 ± 1.8	32.80 ± 1.60	9.97 ± 1.50	5.66 ± 0.85	1.30 ± 0.20
10	1	1	75-76	38.4 ± 1.9	8.40 ± 0.42	2.35 ± 0.35	3.19 ± 0.48	0.85 ± 0.13
10	3	1	75-76	37.9 ± 1.9	3.62 ± 0.18	1.21 ± 0.18	1.27 ± 0.19	0.68 ± 0.10
10	5	3	79-80	36.7 ± 1.8	2.64 ± 0.13	0.86 ± 0.13	0.82 ± 0.12	0.54 ± 0.10
10	7	3	10-11	40.2 ± 2.0	13.30 ± 0.67	3.67 ± 0.55	3.93 ± 0.59	1.40 ± 0.20
10	9	1	63-65	44.0 ± 2.2	14.30 ± 0.72	1.86 ± 0.28	1.74 ± 0.26	1.10 ± 0.20
10	10	1	106-108	39.7 ± 2.0	1.25 ± 0.06	0.49 ± 0.07	0.35 ± 0.05	0.50 ± 0.08
10	10	3	15-17	42.7 ± 2.1	1.57 ± 0.08	0.66 ± 0.10	0.45 ± 0.07	0.44 ± 0.06

TABLE 1 - Continued

Sample Designation				Composition, Wt %				
Hole	Core	Section	Sampled at (cm)	O	Si	Fe	Al	Mg
10	11	3	50-51	38.0 ± 1.9	3.25 ± 0.16	0.95 ± 0.14	0.69 ± 0.10	0.60 ± 0.09
10	12	2	10-12	37.4 ± 1.9	3.79 ± 0.19	1.12 ± 0.17	0.86 ± 0.13	0.83 ± 0.12
10	13	1	100-101	37.6 ± 1.9	3.09 ± 0.15	1.19 ± 0.18	0.61 ± 0.09	0.64 ± 0.10
10	13	3	100-102	38.0 ± 1.9	2.04 ± 0.10	0.74 ± 0.11	0.45 ± 0.07	0.58 ± 0.09
10	14	3	119-120	38.1 ± 1.9	1.70 ± 0.09	0.49 ± 0.07	0.37 ± 0.06	0.50 ± 0.08
10	15	2	119-122	37.4 ± 1.9	2.39 ± 0.12	0.72 ± 0.11	0.57 ± 0.09	0.58 ± 0.09
10	16	2	96-98	37.3 ± 1.9	3.29 ± 0.16	0.86 ± 0.13	0.76 ± 0.11	1.20 ± 0.2
10	17	1	143-144	37.0 ± 1.8	2.00 ± 0.10	0.51 ± 0.08	0.47 ± 0.07	0.61 ± 0.09
10	18	2	99-100	35.8 ± 1.8	1.98 ± 0.10	0.43 ± 0.06	0.49 ± 0.07	0.65 ± 0.10
11	1	1	11-13	40.0 ± 2.0	3.80 ± 0.19	0.96 ± 0.14	1.21 ± 0.18	0.62 ± 0.09
11A	4	2	0-10	38.0 ± 1.9	1.03 ± 0.05	0.61 ± 0.09	0.30 ± 0.05	0.38 ± 0.06
12B	1	1	15-17	45.4 ± 2.3	28.80 ± 1.40	3.77 ± 0.57	7.95 ± 1.19	3.70 ± 0.56
12B	2	2	13-15	45.4 ± 2.3	26.90 ± 1.30	3.66 ± 0.55	5.87 ± 0.88	4.20 ± 0.63
12C	1	1	125-131	39.7 ± 2.0	8.00 ± 0.40	1.60 ± 0.24	2.56 ± 0.38	0.71 ± 0.10
12C	1	4	145-150	38.5 ± 1.9	14.80 ± 0.74	2.67 ± 0.40	4.16 ± 0.62	1.10 ± 0.02
12C	2R	3	145-150	38.8 ± 1.9	11.60 ± 0.58	2.44 ± 0.37	3.55 ± 0.53	1.20 ± 0.02
12C	4	2	18-20	39.4 ± 1.9	10.80 ± 0.54	2.49 ± 0.37	3.59 ± 0.54	1.00 ± 0.02
12C	5	2	78-80	37.3 ± 1.9	26.90 ± 1.30	6.32 ± 0.95	9.78 ± 1.47	1.90 ± 0.03
12C	7	1	16-18	37.0 ± 1.8	26.40 ± 1.30	5.52 ± 0.83	10.88 ± 1.63	2.50 ± 0.04
12C	9	1	124-126	43.3 ± 2.2	26.90 ± 1.30	5.10 ± 0.77	8.09 ± 1.21	2.70 ± 0.04